

ISS-1

Test ID	FNSB EPS
Date	1/8/20
Start Time (hh:mm:ss)	9:40
End Time (hh:mm:ss)	16:40:00
Y (DGM calibration factor)	1.00282
Pre test leak (A cfm @ B in Hg)	< 0.01 @ 12 in Hg
Post test leak (A cfm @ B in Hg)	< 0.01 @ 3 in Hg

Filter Assemblies		
A	FH #	FH-05, FH, 10, FF-23
	Front Filter #	EM-19-00206
	Back Filter #	EM-19-00207
B	FH #	
	Front Filter #	
	Back Filter #	

*Pre and Post leak checks should be 60 seconds in duration

PM Sampling Flow may be 0.15-.25 cfm

Duration (hh:mm:ss)	DGM (ft³)	ΔH (in H ₂ O)
0:00:00	348.527	0.14
0:10:00	350.910	0.14
0:20:00	353.240	0.14
0:30:00	355.550	0.14
0:40:00	357.850	0.14
0:50:00	360.170	0.14
1:00:00	362.450	0.14
1:10:00	364.750	0.14
1:20:00	367.050	0.14
1:30:00	369.335	0.14
1:40:00	371.710	0.14
1:50:00	374.065	0.14
2:00:00	376.390	0.14
2:10:00	378.785	0.14
2:20:00	381.120	0.14
2:30:00	383.585	0.15
2:40:00	386.033	0.15
2:50:00	388.510	0.15
3:00:00	391.020	0.14
3:10:00	393.465	0.15
3:20:00	395.970	0.15
3:30:00	398.440	0.15
3:40:00	400.885	0.15
3:50:00	403.325	0.15
4:00:00	405.720	0.15

Module In Temp (°F)	Module Out Temp (°F)	Vacuum (in Hg)
61	60	1.0
61	60	1.0
62	61	1.0
64	61	1.0
66	63	1.0
68	64	1.0
70	66	1.0
71	68	1.0
72	70	1.0
74	70	1.0
74	72	1.0
75	73	1.0
75	74	1.0
76	75	1.0
76	76	1.0
77	76	1.0
78	76	1.0
78	76	1.0
77	75	1.0
76	75	1.0
75	75	1.0
74	74	1.0
74	74	1.0
73	73	1.0
72	72	1.0

Signature Kelli O'Brien
 Quality Review Kelli O'Brien

Duration (hh:mm:ss)	DGM (ft³)	ΔH (in H₂O)
4:10:00	408.190	0.15
4:20:00	410.620	0.16
4:30:00	413.190	0.16
4:40:00	415.670	0.16
4:50:00	418.165	0.16
5:00:00	420.605	0.16
5:10:00	423.080	0.16
5:20:00	425.630	0.17
5:30:00	428.210	0.17
5:40:00	430.750	0.17
5:50:00	433.310	0.17
6:00:00	435.870	0.17
6:10:00	438.425	0.17
6:20:00	440.985	0.17
6:30:00	443.515	0.17
6:40:00	446.190	0.17
6:50:00	448.590	0.17
7:00:00	451.140	
7:10:00		
7:20:00		
7:30:00		
7:40:00		
7:50:00		
8:00:00		
8:10:00		
8:20:00		
8:30:00		
8:40:00		
8:50:00		
9:00:00		
9:10:00		
9:20:00		
9:30:00		
9:40:00		
9:50:00		
10:00:00		
10:10:00		

[illegible]

Signature _____

Kali OBu